




## Product Specification


The device specification for single lobe LAA

	Code	D Distal Umbrella Diameter (mm)	D1 Proximal Cover Diameter (mm)	Size of LAmbré™ Delivery System
		LT-LAA-1622	16	
	LT-LAA-1824	18	24	8F, 900mm
	LT-LAA-2026	20	26	9F, 900mm
	LT-LAA-2228	22	28	
	LT-LAA-2430	24	30	
	LT-LAA-2632	26	32	10F, 900mm
	LT-LAA-2834	28	34	
	LT-LAA-3036	30	36	
	LT-LAA-3236	32	36	
	LT-LAA-3438	34	38	
	LT-LAA-3640	36	40	


The device specification for multiple lobes LAA

	Code	D Distal Umbrella Diameter (mm)	D1 Proximal Cover Diameter (mm)	Size of LAmbré™ Delivery System
		LT-LAA-1630	16	
	LT-LAA-1832	18	32	10F, 900mm
	LT-LAA-2032	20	32	
	LT-LAA-2234	22	34	
	LT-LAA-2436	24	36	
	LT-LAA-2638	26	38	

LAmbré™ LAA Delivery System with 45°x 30° Double Curves

	Code	Sheath ID (mm)	Sheath OD (mm)	Sheath Length (mm)
	SL8F45×30-900	2.7	3.5	900
	SL9F45×30-900	3.0	3.8	900
	SL10F45×30-900	3.4	4.1	900

LAmbré™ LAA Delivery System with 45° Single Curve

	Code	Sheath ID (mm)	Sheath OD (mm)	Sheath Length (mm)
	SL8F45-900	2.7	3.5	900
	SL9F45-900	3.0	3.8	900
	SL10F45-900	3.4	4.1	900

**LAmbré™**  
LAA Closure System



*Better and Safer Technology for LAA Closure*

## LifeTech Scientific Corporation

Add: Cybio Electronic Building, Langshan 2nd Street, North Area of High-Tech Park,  
Nanshan District, Shenzhen 518057, P. R. China  
Tel: +86-755-86026250 Fax: +86-755-86026251  
Email: marketing@lifetechmed.com  
Web: www.lifetechmed.com

CE  
0086

**LifeTech**  
SCIENTIFIC



LAA Closure System

## A Novel Left Atrial Appendage (LAA) Closure Device

### Easy to Use and Safe

#### Innovative umbrella design

- Close LAA regardless of its anatomy
- Fully retrievable and repositionable

#### Patented anchor design for better device fixation

- Small distal hooks to engage LAA walls
- Round U-shaped Ends to trap trabeculations
- Over-sized umbrella to stent against LAA

#### 8-10 Fr. small delivery sheath


#### The delivery system is available in

- double curves (45° X 30°) of distal tip
- single curve (45°) of distal tip




## Global Clinical Trials Data Demonstrated LAMBRE™ Safe and Easy for LAA Closure

### High implant success<sup>[1]</sup>



**99.6%**  
(n=218)  
Global Study  
Implant  
Success

### Successful implantation regardless of LAA anatomies



Could include patients who may be rejected by other devices due to short LAA depth.

**100%**  
(n=218)  
Patient Inclusion  
Regardless of  
LAA Anatomies

### Stable fixation

No device embolization or migration.

**0.00%**  
(n=218)  
Device  
Embolization

[1] Implant success: stable placement within LAA and immediately residual flow less than 5mm.

